

 Dr. Arabin dare to care	<u>Dr. Arabin GmbH & Co. KG</u> Declaration of Conformity	Page: 1 from 1 Rev. / Date: 20 / 02.09.2021
---	--	---

Declaration of Conformity

Company:

Dr. Arabin GmbH & Co. KG
Alfred Herrhausen St. 44
58455 Witten

We declare under our sole responsibility that the medical device

- Cerclage pessary type A - (CA xx/xx/xx)
- Cerclage pessary type ASQ - (ASQ xx/xx/xx)

Directive	93/42/EEC
Classification	Class IIb
Applied rules	Rule 5

corresponds.

It fulfills the Essential Requirements according to Annex I of Directive 93/42 EEC.

Conformity assessment procedures:

Class IIb Annex II without section 4


Certificates:

QS - 1340 Certificate No. 1340GB410191127 (valid until May 27 2024)

Notified body:

MEDCERT GmbH, Pilatuspool 2, 20355 Hamburg, Germany
Identification number: 0482

This declaration of conformity is valid until the expiry of the above certificates or if a change is made to the product.

Department	Checked on	Released by
Management	02.09.2021	
QM representative	02.09.2021	M. Prior